

**Annex of the certificate (Page 1/38)
Accreditation Scope**



Kalibrasyon
TS EN ISO/IEC 17025
AB-0016-K

METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ.
Kalibrasyon Laboratuvarı
Accreditation Nr: AB-0016-K
Revision Nr: 015 Date: 19.11.2019

As a Calibration Laboratory

Address :

Çavuşoğlu Mah. Barbaros Hayrettin Paşa Cad.
No:16-18-21 KARTAL 81430
İSTANBUL/TÜRKİYE

Phone : 0 216 374 99 24

Fax : 0 216 374 99 28

E-mail : metkal@metkal.com.tr

Website : www.metkal.com.tr

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
GAS MIXTURES Gas Analyzer Carbon monoxide	98,9 µmol/mol ≤ C ≤ 100,9 µmol/mol	By using traceable gas mixtures of N ₂ matrix	% 2,4	C : Gas concentration Comparison method with certified gas mixtures used as working standard
Nitrogen monoxide	804 µmol/mol ≤ C ≤ 820 µmol/mol	By using traceable gas mixtures of N ₂ matrix	% 2,3	C : Gas concentration Comparison method with certified gas mixtures used as working standard
Oxygen	20,9 cmol/mol ≤ C ≤ 21,3 cmol/mol	By using traceable gas mixtures of N ₂ matrix	% 2,3	C : Gas concentration Comparison method with certified gas mixtures used as working standard
Sulfur dioxide	966 µmol/mol ≤ C ≤ 985 µmol/mol	By using traceable gas mixtures of N ₂ matrix	% 2,3	C : Gas concentration Comparison method with certified gas mixtures used as working standard
OPTICAL QUANTITIES Illuminance Lux Meter, Light Meter, Illuminance Meter	0,1 lx ≤ E _v ≤ 5000 lx	2856 K	% 0,96	E _v : Illuminance Reference method: CIE S023 / ISO19476

Annex of the certificate (Page 2/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Radiometry UV-A Detector, UV-A Radiometry	$19 \mu\text{W}/\text{cm}^2 \leq E_e \leq 20000 \mu\text{W}/\text{cm}^2$	Meters used within UV-A range	% 4,3	E_e : Radiometry Reference method: CIE 202
ACOUSTIC Sound Pressure Response Level Sound Level Meter	70 dB-130 dB	$31,5 \text{ Hz} \leq f \leq 16000 \text{ Hz}$	0,3 dB	f : Frequency Calibration by comparison method with multifunction acoustic calibrator
ELECTRICAL DC Voltage ($\leq 1100 \text{ V}$) DC Voltage Source	$0,2 \text{ mV} \leq U < 200 \text{ mV}$ $0,2 \text{ V} \leq U < 2 \text{ V}$ $2 \text{ V} \leq U < 20 \text{ V}$ $20 \text{ V} \leq U < 200 \text{ V}$ $200 \text{ V} \leq U < 1000 \text{ V}$		$6,2 \cdot 10^{-6} \cdot U + 0,12 \mu\text{V}$ $2,1 \cdot 10^{-6} \cdot U + 6,0 \mu\text{V}$ $2,2 \cdot 10^{-6} \cdot U + 60 \mu\text{V}$ $6,6 \cdot 10^{-6} \cdot U + 0,04 \text{ mV}$ $9,0 \cdot 10^{-6} \cdot U$	U : Voltage
DC High Voltage ($> 1100 \text{ V}$) DC Voltage Source	$1 \text{ kV} \leq U < 40 \text{ kV}$		% 1,0	U : Voltage By means of high voltage probe
DC Voltage ($\leq 1100 \text{ V}$) DC Voltage Measuring Instrument, Multimeter, Voltmeter	$0 \text{ mV} \leq U < 329,99 \text{ mV}$ $0,33 \text{ V} \leq U < 3,299999 \text{ V}$ $3,3 \text{ V} \leq U < 32,99999 \text{ V}$ $33 \text{ V} \leq U < 329,9999 \text{ V}$ $330 \text{ V} \leq U < 1000 \text{ V}$		$5,1 \cdot 10^{-5} \cdot U + 2,4 \mu\text{V}$ $4,5 \cdot 10^{-5} \cdot U + 3,5 \mu\text{V}$ $4,5 \cdot 10^{-5} \cdot U + 37 \mu\text{V}$ $4,8 \cdot 10^{-5} \cdot U + 0,36 \text{ mV}$ $4,9 \cdot 10^{-5} \cdot U + 1,0 \text{ mV}$	U : Voltage

Annex of the certificate (Page 3/38)


Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
DC Current	0,2 µA ≤ I < 200 µA		1,7 · 10 ⁻⁵ · I + 0,41 nA	/: Current
DC Current Source, Calibrator	0,2 mA ≤ I < 2 mA		1,8 · 10 ⁻⁵ · I + 3,56 nA	
	2 mA ≤ I < 20 mA		1,9 · 10 ⁻⁵ · I + 37 nA	
	20 mA ≤ I < 200 mA		5,0 · 10 ⁻⁵ · I + 0,9 µA	
	0,2 A ≤ I < 2 A		1,9 · 10 ⁻⁴ · I + 17 µA	
	2 A ≤ I < 20 A		4,1 · 10 ⁻⁴ · I + 0,45 mA	
DC Current	10 A ≤ I < 100 A		1,4 · 10 ⁻² · I + 0,17 A	/: Current
DC Current Source, Calibrator	100 A ≤ I ≤ 900 A		1,2 · 10 ⁻³ · I + 0,53 A	
DC Current	0 µA ≤ I < 329,999 µA		1,2 · 10 ⁻⁴ · I + 19 nA	/: Current
DC Current Measuring Instrument, Multimeter: DC Current, Ammeter, Clamp Meter (Current Clamp)	0,33 mA ≤ I < 3,29999 mA		8,6 · 10 ⁻⁵ · I + 0,06 µA	
	3,3 mA ≤ I < 32,9999 mA		1,0 · 10 ⁻⁴ · I + 0,20 µA	
	33 mA ≤ I < 329,999 mA		8,6 · 10 ⁻⁵ · I + 2,0 µA	
	0,33 A ≤ I < 2,99999 A		2,9 · 10 ⁻⁴ · I + 50 µA	
	3 A ≤ I < 10,9999 A		4,7 · 10 ⁻⁴ · I + 0,38 mA	
	11 A ≤ I < 20,5 A		6,4 · 10 ⁻⁴ · I + 5,5 mA	

Annex of the certificate (Page 4/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
DC Current DC Current Measuring Instrument, Multimeter: DC Current, Ammeter, Clamp Meter (Current Clamp)	10 A ≤ I < 16,5 A		$2,9 \cdot 10^{-3} \cdot I + 2,3 \text{ mA}$	/: Current 50 turn coil Toroidal type meter
	16,5 A ≤ I < 150 A		$2,9 \cdot 10^{-3} \cdot I + 17 \text{ mA}$	
	150 A ≤ I < 550 A		$3,0 \cdot 10^{-3} \cdot I + 70 \text{ mA}$	
	550 A ≤ I < 1000 A		$3,0 \cdot 10^{-3} \cdot I + 68 \text{ mA}$	
DC Current DC Current Measuring Instrument, Multimeter: DC Current, Ammeter, Clamp Meter (Current Clamp)	10 A ≤ I < 16,5 A		$5,8 \cdot 10^{-3} \cdot I + 23 \text{ mA}$	/: Current 50 turn coil Non-toroidal type meter
	16,5 A ≤ I < 150 A		$5,8 \cdot 10^{-3} \cdot I + 0,16 \text{ A}$	
	150 A ≤ I < 550 A		$5,8 \cdot 10^{-3} \cdot I + 0,58 \text{ A}$	
	550 A ≤ I < 1000 A		$5,9 \cdot 10^{-3} \cdot I + 0,57 \text{ A}$	
DC Power Power meter, Wattmeter	10,9 W ≤ P < 726 W	33 V - 330 V 0,33 A - 2,2 A	$3,3 \cdot 10^{-4} \cdot P + 0,9 \text{ mW}$	P: Power
	72,6 W ≤ P < 3630 W	33 V - 330 V 2,2 A - 11 A	$8,4 \cdot 10^{-4} \cdot P - 37 \text{ mW}$	
	220 W ≤ P < 11000 W	100 V - 1000 V 2,2 A - 11 A	$5,1 \cdot 10^{-4} \cdot P - 43 \text{ mW}$	
	1100 W ≤ P < 20500 W	100 V - 1000 V 11 A - 20,5 A	$4,9 \cdot 10^{-4} \cdot P + 24 \text{ mW}$	
DC Power Power meter, Wattmeter	0,54 kW ≤ P < 36,3 kW	33 V - 330 V 16,5 A - 110 A	$3,0 \cdot 10^{-3} \cdot P + 0,8 \text{ W}$	P: Power 50 turn coil Toroidal type meter
	3,6 kW ≤ P < 181,5 kW	33 V - 330 V 110 A - 550 A	$3,0 \cdot 10^{-3} \cdot P + 0,2 \text{ W}$	
	11 kW ≤ P < 1000 kW	100 V - 1000 V 550 A - 1000 A	$3,0 \cdot 10^{-3} \cdot P + 5,1 \text{ W}$	

Annex of the certificate (Page 5/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
DC Power Power meter, Wattmeter	$0,54 \text{ kW} \leq P < 36,3 \text{ kW}$	33 V - 330 V 16,5 A - 110 A	$7,7 \cdot 10^{-3} \cdot P + 6,6 \text{ W}$	P: Power 50 turn coil Non-toroidal type meter
	$3,6 \text{ kW} \leq P < 181,5 \text{ kW}$	33 V - 330 V 110 A - 550 A	$6,1 \cdot 10^{-3} \cdot P + 4,3 \text{ W}$	
	$11 \text{ kW} \leq P < 1000 \text{ kW}$	100 V - 1000 V 550 A - 1000 A	$6,4 \cdot 10^{-3} \cdot P + 9,9 \text{ W}$	
AC Voltage ($\leq 1100 \text{ V}$) AC Voltage Source	$0,2 \text{ mV} \leq U < 200 \text{ mV}$	10 Hz - 10 kHz	$1,6 \cdot 10^{-4} \cdot U + 4,1 \mu\text{V}$	U: Voltage
		10 kHz - 30 kHz	$3,6 \cdot 10^{-4} \cdot U + 8,2 \mu\text{V}$	
		30 kHz - 100 kHz	$7,9 \cdot 10^{-4} \cdot U + 20 \mu\text{V}$	
	$0,2 \text{ V} \leq U < 2 \text{ V}$	10 Hz - 10 kHz	$1,3 \cdot 10^{-4} \cdot U + 0,02 \text{ mV}$	
		10 kHz - 30 kHz	$2,3 \cdot 10^{-4} \cdot U + 0,04 \text{ mV}$	
		30 kHz - 100 kHz	$5,9 \cdot 10^{-4} \cdot U + 0,20 \text{ mV}$	
AC Voltage ($\leq 1100 \text{ V}$) AC Voltage Source	$2 \text{ V} \leq U < 20 \text{ V}$	10 Hz - 10 kHz	$1,3 \cdot 10^{-4} \cdot U + 0,20 \text{ mV}$	U: Voltage
		10 kHz - 30 kHz	$2,3 \cdot 10^{-4} \cdot U + 0,41 \text{ mV}$	
		30 kHz - 100 kHz	$5,9 \cdot 10^{-4} \cdot U + 2,0 \text{ mV}$	
	$20 \text{ V} \leq U < 200 \text{ V}$	10 Hz - 10 kHz	$1,3 \cdot 10^{-4} \cdot U + 2,0 \text{ mV}$	
		10 kHz - 30 kHz	$2,3 \cdot 10^{-4} \cdot U + 4,0 \text{ mV}$	
		30 kHz - 100 kHz	$5,9 \cdot 10^{-4} \cdot U + 20 \text{ mV}$	
AC Voltage ($\leq 1100 \text{ V}$) AC Voltage Source	$200 \text{ V} \leq U < 1000 \text{ V}$	10 Hz - 10 kHz	$1,5 \cdot 10^{-4} \cdot U + 0,02 \text{ V}$	U: Voltage
		10 kHz - 30 kHz	$2,9 \cdot 10^{-4} \cdot U - 0,01 \text{ V}$	

Annex of the certificate (Page 6/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
AC Voltage (≤ 1100 V) AC Voltage Source	$30 \text{ mV} \leq U < 30 \text{ V}$	100 kHz - 60 MHz	% 4	U: Voltage Measuring by oscilloscope
AC High Voltage (> 1100 V) AC Voltage Source	$1 \text{ kV} \leq U < 28 \text{ kV}$	50 Hz	% 1,0	U: Voltage By means of high voltage probe
AC Voltage (≤ 1100 V) AC Voltage Measuring Instrument, Multimeter, Voltmeter	$1 \text{ mV} \leq U < 32,999 \text{ mV}$	10 Hz - 45 Hz 45 Hz - 20 kHz 20kHz - 100 kHz 100 kHz - 500 kHz	$1,2 \cdot 10^{-3} \cdot U + 16 \mu\text{V}$ $1,2 \cdot 10^{-3} \cdot U + 16 \mu\text{V}$ $2,8 \cdot 10^{-3} \cdot U + 26 \mu\text{V}$ $7,9 \cdot 10^{-3} \cdot U + 42 \mu\text{V}$	U: Voltage
AC Voltage (≤ 1100 V) AC Voltage Measuring Instrument, Multimeter, Voltmeter	$33 \text{ mV} \leq U < 329,999 \text{ mV}$	10 Hz - 45 Hz 45 Hz - 10 kHz 10 kHz - 50 kHz 50 kHz - 100 kHz 100 kHz - 500 kHz	$3,9 \cdot 10^{-4} \cdot U + 16 \mu\text{V}$ $2,4 \cdot 10^{-4} \cdot U + 16 \mu\text{V}$ $7,9 \cdot 10^{-4} \cdot U + 31 \mu\text{V}$ $1,8 \cdot 10^{-3} \cdot U + 0,13 \text{ mV}$ $3,9 \cdot 10^{-3} \cdot U + 0,26 \text{ mV}$	U: Voltage
AC Voltage (≤ 1100 V) AC Voltage Measuring Instrument, Multimeter, Voltmeter	$0,33 \text{ V} \leq U < 3,29999 \text{ V}$	10 Hz - 45 Hz 45 Hz - 10 kHz 10 kHz - 50 kHz 50 kHz - 100 kHz 100 kHz - 500 kHz	$4,0 \cdot 10^{-4} \cdot U + 0,05 \text{ mV}$ $2,4 \cdot 10^{-4} \cdot U + 0,47 \text{ mV}$ $7,9 \cdot 10^{-4} \cdot U + 0,04 \text{ mV}$ $1,8 \cdot 10^{-3} \cdot U + 0,2 \text{ mV}$ $4,0 \cdot 10^{-3} \cdot U + 0,7 \text{ mV}$	U: Voltage

Annex of the certificate (Page 7/38)


Accreditation Scope

 <p>TÜRKAK Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
AC Voltage (≤ 1100 V) AC Voltage Measuring Instrument, Multimeter, Voltmeter	$3,3 \text{ V} \leq U < 32,9999 \text{ V}$	10 Hz - 45 Hz	$4,0 \cdot 10^{-4} \cdot U + 0,6 \text{ mV}$	U: Voltage
		45 Hz - 10 kHz	$2,4 \cdot 10^{-4} \cdot U + 0,46 \text{ mV}$	
		10 kHz - 20 kHz	$5,5 \cdot 10^{-4} \cdot U + 0,5 \text{ mV}$	
		20 kHz - 50 kHz	$7,9 \cdot 10^{-4} \cdot U + 0,45 \text{ mV}$	
		50 kHz - 100 kHz	$1,8 \cdot 10^{-3} \cdot U + 1,5 \text{ mV}$	
AC Voltage (≤ 1100 V) AC Voltage Measuring Instrument, Multimeter, Voltmeter	$33 \text{ V} \leq U < 329,999 \text{ V}$	45 Hz - 1 kHz	$4,0 \cdot 10^{-4} \cdot U + 2,2 \text{ mV}$	U: Voltage
		1 kHz - 10 kHz	$6,3 \cdot 10^{-4} \cdot U + 7,0 \text{ mV}$	
		10 kHz - 20 kHz	$7,1 \cdot 10^{-4} \cdot U + 7,0 \text{ mV}$	
AC Voltage (≤ 1100 V) AC Voltage Measuring Instrument, Multimeter, Voltmeter	$330 \text{ V} \leq U < 1020 \text{ V}$	45 Hz - 1 kHz	$4,4 \cdot 10^{-4} \cdot U + 13 \text{ mV}$	U: Voltage
		1 kHz - 10 kHz	$7,1 \cdot 10^{-4} \cdot U + 15 \text{ mV}$	
AC Current AC Current Source, Calibrator	$20 \mu\text{A} \leq I < 200 \mu\text{A}$	10 Hz - 10 kHz	$5,9 \cdot 10^{-4} \cdot I + 20 \text{ nA}$	I: Current
	$0,2 \text{ mA} \leq I < 2 \text{ mA}$	10 Hz - 10 kHz	$3,3 \cdot 10^{-4} \cdot I + 0,2 \mu\text{A}$	
	$2 \text{ mA} \leq I < 20 \text{ mA}$	10 Hz - 10 kHz	$3,4 \cdot 10^{-4} \cdot I + 2,0 \mu\text{A}$	
	$20 \text{ mA} \leq I < 200 \text{ mA}$	10 Hz - 10 kHz	$3,3 \cdot 10^{-4} \cdot I + 20 \mu\text{A}$	
AC Current AC Current Source, Calibrator	$0,2 \text{ A} \leq I < 2 \text{ A}$	10 Hz - 10 kHz	$7,7 \cdot 10^{-4} \cdot I + 0,2 \text{ mA}$	I: Current
	$2 \text{ A} \leq I < 20 \text{ A}$	10 Hz - 10 kHz	$8,6 \cdot 10^{-4} \cdot I + 2,0 \text{ mA}$	

Annex of the certificate (Page 8/38)

Accreditation Scope

 Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ.</p> <p>Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K</p> <p>Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
AC Current	10 A ≤ I < 100 A	50 Hz	$1,3 \cdot 10^{-2} \cdot I + 0,18 \text{ A}$	/: Current
AC Current	100 A ≤ I < 900 A	50 Hz	$1,2 \cdot 10^{-2} \cdot I + 0,51 \text{ A}$	
Source, Calibrator	250 A ≤ I < 2500 A	50 Hz	$9,2 \cdot 10^{-4} \cdot I + 0,57 \text{ A}$	
AC Current	29 µA ≤ I < 329,99 µA	10 Hz - 1 kHz	$1,6 \cdot 10^{-3} \cdot I + 0,08 \text{ µA}$	/: Current
AC Current		1 kHz - 10 kHz	$6,2 \cdot 10^{-3} \cdot I + 0,16 \text{ µA}$	
Measuring		10 kHz - 30 kHz	$1,2 \cdot 10^{-2} \cdot I + 0,31 \text{ µA}$	
Instrument, Multimeter: AC				
Current, AC				
Ammeter, Clamp Meter				
(Current Clamp)				
AC Current	0,33 mA ≤ I < 3,29999 mA	10 Hz - 20 Hz	$1,6 \cdot 10^{-3} \cdot I + 0,12 \text{ µA}$	/: Current
AC Current		20 Hz - 1 kHz	$1,0 \cdot 10^{-3} \cdot I + 0,12 \text{ µA}$	
Measuring		1 kHz - 5 kHz	$1,6 \cdot 10^{-3} \cdot I + 0,16 \text{ µA}$	
Instrument, Multimeter: AC		5 kHz - 10 kHz	$3,9 \cdot 10^{-3} \cdot I + 0,23 \text{ µA}$	
Current, AC		10 kHz - 30 kHz	$7,8 \cdot 10^{-3} \cdot I + 0,47 \text{ µA}$	
Ammeter, Clamp Meter				
(Current Clamp)				
AC Current	3,3 mA ≤ I < 32,9999 mA	10 Hz - 20 Hz	$1,4 \cdot 10^{-3} \cdot I + 1,6 \text{ µA}$	/: Current
AC Current		20 Hz - 1 kHz	$3,6 \cdot 10^{-4} \cdot I + 1,5 \text{ µA}$	
Measuring		1 kHz - 5 kHz	$6,5 \cdot 10^{-4} \cdot I + 1,6 \text{ µA}$	
Instrument, Multimeter: AC		5 kHz - 10 kHz	$1,6 \cdot 10^{-3} \cdot I + 2,4 \text{ µA}$	
Current, AC		10 kHz - 30 kHz	$6,9 \cdot 10^{-3} \cdot I + 3,5 \text{ µA}$	
Ammeter, Clamp Meter				
(Current Clamp)				

Annex of the certificate (Page 9/38)

Accreditation Scope

 Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ.</p> <p>Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K</p> <p>Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
AC Current AC Current Measuring Instrument, Multimeter: AC Current, AC Ammeter, Clamp Meter (Current Clamp)	$33 \text{ mA} \leq I < 329,999 \text{ mA}$	10 Hz - 20 Hz	$1,4 \cdot 10^{-3} \cdot I + 15 \text{ } \mu\text{A}$	/: Current
		20 Hz - 1 kHz	$7,2 \cdot 10^{-4} \cdot I + 15 \text{ } \mu\text{A}$	
		1 kHz - 5 kHz	$3,6 \cdot 10^{-4} \cdot I + 15 \text{ } \mu\text{A}$	
		5 kHz - 10 kHz	$8,0 \cdot 10^{-4} \cdot I + 39 \text{ } \mu\text{A}$	
		10 kHz - 30 kHz	$3,1 \cdot 10^{-3} \cdot I + 0,16 \text{ mA}$	
AC Current AC Current Measuring Instrument, Multimeter: AC Current, AC Ammeter, Clamp Meter (Current Clamp)	$0,33 \text{ A} \leq I < 2,99999 \text{ A}$	10 Hz - 45 Hz	$1,4 \cdot 10^{-3} \cdot I + 0,08 \text{ mA}$	/: Current
		45 Hz - 1 kHz	$4,7 \cdot 10^{-4} \cdot I + 0,09 \text{ mA}$	
		1 kHz - 5 kHz	$4,7 \cdot 10^{-3} \cdot I + 0,8 \text{ mA}$	
		5 kHz - 10 kHz	$1,9 \cdot 10^{-2} \cdot I + 3,9 \text{ mA}$	
AC Current AC Current Measuring Instrument, Multimeter: AC Current, AC Ammeter, Clamp Meter (Current Clamp)	$3 \text{ A} \leq I < 10,9999 \text{ A}$	45 Hz - 1 kHz	$1,0 \cdot 10^{-3} \cdot I + 0,9 \text{ mA}$	/: Current
		1 kHz - 5 kHz	$2,3 \cdot 10^{-2} \cdot I + 1,5 \text{ mA}$	
	$11 \text{ A} \leq I < 20,5 \text{ A}$	45 Hz - 1 kHz	$1,2 \cdot 10^{-3} \cdot I + 3,9 \text{ mA}$	
		1 kHz - 5 kHz	$2,3 \cdot 10^{-2} \cdot I + 3,9 \text{ mA}$	
AC Current AC Current Measuring Instrument, Multimeter: AC Current, AC Ammeter, Clamp Meter (Current Clamp), Weld-Tester	$10 \text{ A} \leq I < 16,5 \text{ A}$	45 Hz - 65 Hz	$3,3 \cdot 10^{-3} \cdot I + 3,5 \text{ mA}$	/: Current 50 turn coil Toroidal type meter
		65 Hz - 440 Hz	$9,2 \cdot 10^{-3} \cdot I + 3,5 \text{ mA}$	
	$16,5 \text{ A} \leq I < 150 \text{ A}$	45 Hz - 65 Hz	$9,3 \cdot 10^{-4} \cdot I + 0,38 \text{ A}$	
		65 Hz - 440 Hz	$9,2 \cdot 10^{-3} \cdot I + 0,03 \text{ A}$	
	$150 \text{ A} \leq I < 550 \text{ A}$	45 Hz - 65 Hz	$3,4 \cdot 10^{-3} \cdot I + 0,13 \text{ A}$	
		65 Hz - 440 Hz	$9,2 \cdot 10^{-3} \cdot I + 0,13 \text{ A}$	
	$550 \text{ A} \leq I < 1000 \text{ A}$	45 Hz - 65 Hz	$3,4 \cdot 10^{-3} \cdot I + 0,16 \text{ A}$	
		65 Hz - 440 Hz	$9,2 \cdot 10^{-3} \cdot I + 0,14 \text{ A}$	

Annex of the certificate (Page 10/38)

Accreditation Scope

 Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ.</p> <p>Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K</p> <p>Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
AC Current AC Current Measuring Instrument, Multimeter: AC Current, AC Ammeter, Clamp Meter (Current Clamp), Weld-Tester	10 A ≤ I < 16,5 A	45 Hz - 65 Hz 65 Hz - 440 Hz	6,5 · 10 ⁻³ · I + 35 mA 1,2 · 10 ⁻² · I + 35 mA	I: Current 50 turn coil Non-toroidal type meter
	16,5 A ≤ I < 150 A	45 Hz - 65 Hz 65 Hz - 440 Hz	6,5 · 10 ⁻³ · I + 0,29 A 1,2 · 10 ⁻² · I + 0,29 A	
	150 A ≤ I < 550 A	45 Hz - 65 Hz 65 Hz - 440 Hz	6,5 · 10 ⁻³ · I + 1,0 A 1,2 · 10 ⁻² · I + 1,0 A	
	550 A ≤ I < 1000 A	45 Hz - 65 Hz 65 Hz - 440 Hz	6,6 · 10 ⁻³ · I + 1,1 A 1,2 · 10 ⁻² · I + 1,1 A	
AC Current AC Current Measuring Instrument, Multimeter: AC Current, AC Ammeter, Clamp Meter (Current Clamp), Weld-Tester	550 A ≤ I < 2000 A	50 Hz	% 2,4	I: Current
	2 kA ≤ I < 15 kA	50 Hz	% 2,4	
AC Power and Energy Active Power: Single Phase, Three Phase Power meter, Wattmeter	10,9 W ≤ P < 726 W	U: 33 V - 330 V I: 0,33 A - 2,2 A F: 45 Hz - 65 Hz	6,7 · 10 ⁻⁴ · P + 1,2 mW	P: Power Only for sine wave
	72,6 W ≤ P < 3630 W	U: 33 V - 330 V I: 2,2 A - 11 A F: 45 Hz - 65 Hz	7,7 · 10 ⁻⁴ · P - 5,8 mW	
	220 W ≤ P < 11000 W	U: 100 V - 1000 V I: 2,2 A - 11 A F: 45 Hz - 65 Hz	7,4 · 10 ⁻⁴ · P - 4,8 mW	
	11000 W ≤ P < 20500 W	U: 100 V - 1000 V I: 11 A - 20,5 A F: 45 Hz - 65 Hz	1,2 · 10 ⁻³ · P + 0,18 mW	

Annex of the certificate (Page 11/38)


Accreditation Scope

 Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ.</p> <p>Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K</p> <p>Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
AC Power and Energy Active Power: Single Phase, Three Phase Power meter, Wattmeter	$0,54 \text{ kW} \leq P < 36,3 \text{ kW}$	U: 33 V - 330 V I: 16,5 A - 110 A F: 45 Hz - 65 Hz	$3,6 \cdot 10^{-3} \cdot P + 0,2 \text{ W}$	P: Power Only for sine wave Toroidal type meter
	$3,6 \text{ kW} \leq P < 181,5 \text{ kW}$	U: 33 V - 330 V I: 110 A - 550 A F: 45 Hz - 65 Hz	$4,0 \cdot 10^{-3} \cdot P - 1,3 \text{ W}$	
	$11 \text{ kW} \leq P < 550 \text{ kW}$	U: 100 V - 1000 V I: 110 A - 550 A F: 45 Hz - 65 Hz	$3,3 \cdot 10^{-3} \cdot P + 4,1 \text{ W}$	
AC Power and Energy Active Power: Single Phase, Three Phase Power meter, Wattmeter	$0,54 \text{ kW} \leq P < 36,3 \text{ kW}$	U: 33 V - 330 V I: 16,5 A - 110 A F: 45 Hz - 65 Hz	$9,2 \cdot 10^{-3} \cdot P - 0,2 \text{ W}$	P: Power Only for sine wave Non-toroidal type meter
	$3,6 \text{ kW} \leq P < 181,5 \text{ kW}$	U: 33 V - 330 V I: 110 A - 550 A F: 45 Hz - 65 Hz	$8,6 \cdot 10^{-3} \cdot P + 2,0 \text{ W}$	
	$11 \text{ kW} \leq P < 550 \text{ kW}$	U: 100 V - 1000 V I: 110 A - 550 A F: 45 Hz - 65 Hz	$7,5 \cdot 10^{-3} \cdot P + 19 \text{ W}$	
AC Power and Energy Active Power: Single Phase, Three Phase Power meter, Wattmeter	$0,5 \text{ kW} \leq P < 10 \text{ kW}$	50 Hz	$2,4 \cdot 10^{-2} \cdot P + 14 \text{ W}$	P: Power Direct measurement Calibration on-site
	$10 \text{ kW} \leq P < 100 \text{ kW}$	50 Hz	$2,4 \cdot 10^{-2} \cdot P + 34 \text{ W}$	
	$100 \text{ kW} \leq P < 121 \text{ kW}$	50 Hz	$2,3 \cdot 10^{-2} \cdot P + 0,4 \text{ kW}$	

Annex of the certificate (Page 12/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
DC Resistance DC Resistance Measuring Instrument, Multimeter: Resistance, Ohmeter, Wheatstone Bridge, Insulation Tester	100 $\mu\Omega$ 1 m Ω 10 m Ω 100 m Ω 1 Ω 10 Ω		$8,8 \cdot 10^{-2} \mu\Omega$ $8,7 \cdot 10^{-4} \text{ m}\Omega$ $8,7 \cdot 10^{-3} \text{ m}\Omega$ $8,7 \cdot 10^{-2} \text{ m}\Omega$ $9,0 \cdot 10^{-4} \Omega$ $9,0 \cdot 10^{-3} \Omega$	4-wire Direct measurement
DC Resistance DC Resistance Measuring Instrument, Multimeter: Resistance, Ohmeter, Wheatstone Bridge, Insulation Tester	$0 \Omega \leq R < 10,999 \Omega$ $11 \Omega \leq R < 32,999 \Omega$ $33 \Omega \leq R < 109,999 \Omega$ $110 \Omega \leq R < 329,999 \Omega$		$7,5 \cdot 10^{-5} \cdot R + 10 \text{ m}\Omega$ $8,7 \cdot 10^{-5} \cdot R + 13 \text{ m}\Omega$ $6,8 \cdot 10^{-5} \cdot R + 13 \text{ m}\Omega$ $7,1 \cdot 10^{-5} \cdot R + 16 \text{ m}\Omega$	R: Resistance 4-wire
DC Resistance DC Resistance Measuring Instrument, Multimeter: Resistance, Ohmeter, Wheatstone Bridge, Insulation Tester	$0,33 \text{ k}\Omega \leq R < 1,09999 \text{ k}\Omega$ $1,1 \text{ k}\Omega \leq R < 3,29999 \text{ k}\Omega$ $3,3 \text{ k}\Omega \leq R < 10,9999 \text{ k}\Omega$ $11 \text{ k}\Omega \leq R < 32,9999 \text{ k}\Omega$ $33 \text{ k}\Omega \leq R < 109,999 \text{ k}\Omega$		$5,2 \cdot 10^{-5} \cdot R + 58 \text{ m}\Omega$ $7,1 \cdot 10^{-5} \cdot R + 0,16 \Omega$ $4,9 \cdot 10^{-5} \cdot R + 0,55 \Omega$ $7,0 \cdot 10^{-5} \cdot R + 0,92 \Omega$ $6,4 \cdot 10^{-5} \cdot R + 5,2 \Omega$	R: Resistance 4-wire

Annex of the certificate (Page 13/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
DC Resistance	$110 \text{ k}\Omega \leq R < 329,999 \text{ k}\Omega$		$1,0 \cdot 10^{-4} \cdot R + 8,5 \Omega$	R: Resistance 2-wire
DC Resistance				
Measuring Instrument, Multimeter:	$330 \text{ k}\Omega \leq R < 1,09999 \text{ M}\Omega$		$1,0 \cdot 10^{-4} \cdot R + 47 \Omega$	
Resistance, Ohm meter, Wheatstone Bridge, Insulation Tester	$1,1 \text{ M}\Omega \leq R < 3,29999 \text{ M}\Omega$		$1,6 \cdot 10^{-4} \cdot R + 0,08 \text{ k}\Omega$	
	$3,3 \text{ M}\Omega \leq R < 10,9999 \text{ M}\Omega$		$4,7 \cdot 10^{-4} \cdot R + 0,35 \text{ k}\Omega$	
	$11 \text{ M}\Omega \leq R < 32,9999 \text{ M}\Omega$		$8,0 \cdot 10^{-4} \cdot R + 1,8 \text{ k}\Omega$	
	$33 \text{ M}\Omega \leq R < 109,999 \text{ M}\Omega$		$4,0 \cdot 10^{-3} \cdot R + 2,7 \text{ k}\Omega$	
	$110 \text{ M}\Omega \leq R < 329,999 \text{ M}\Omega$		$4,1 \cdot 10^{-3} \cdot R + 0,06 \text{ M}\Omega$	
	$330 \text{ M}\Omega \leq R < 1100,00 \text{ M}\Omega$		$1,3 \cdot 10^{-2} \cdot R + 0,12 \text{ M}\Omega$	
DC Resistance	$1 \Omega \leq R < 100 \Omega$		$3,2 \cdot 10^{-4} \cdot R + 0,9 \text{ m}\Omega$	R: Resistance Substitution Method
DC Resistance				
Measuring Instrument, Multimeter:	$0,1 \text{ k}\Omega \leq R < 1 \text{ k}\Omega$		$3,4 \cdot 10^{-4} \cdot R + 1,1 \text{ m}\Omega$	
Resistance, Ohm meter, Wheatstone Bridge, Insulation Tester	$1 \text{ k}\Omega \leq R < 10 \text{ k}\Omega$		$3,4 \cdot 10^{-4} \cdot R + 5,7 \text{ m}\Omega$	
	$10 \text{ k}\Omega \leq R < 100 \text{ k}\Omega$		$3,4 \cdot 10^{-4} \cdot R + 0,06 \Omega$	
	$100 \text{ k}\Omega \leq R < 1 \text{ M}\Omega$		$3,4 \cdot 10^{-4} \cdot R + 0,6 \Omega$	

Annex of the certificate (Page 14/38)


Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
DC Resistance	10 MΩ	≤ 5 kV	$8,0 \cdot 10^{-3} \cdot R$	R: Resistance Substitution Method
DC Resistance	100 MΩ		$8,0 \cdot 10^{-3} \cdot R$	
Measuring Instrument,	1 GΩ		$8,1 \cdot 10^{-3} \cdot R$	
Multimeter:	10 GΩ		$8,5 \cdot 10^{-3} \cdot R$	
Resistance, Ohmeter, Wheatstone Bridge, Insulation Tester				
DC Resistance		Within (23±3)°C		4-wire measurement comparison with reference
DC Resistance Standard,	100 μΩ	10 A	$4,9 \cdot 10^{-2} \mu\Omega$	
Calibrator:	1 mΩ	10 A	$7,5 \cdot 10^{-2} \mu\Omega$	
Resistance,	10 mΩ	1 A	$9,2 \cdot 10^{-1} \text{ m}\Omega$	
DC Current Shunt,	100 mΩ	1 A	$6,3 \cdot 10^{-3} \text{ m}\Omega$	
Resistance Box	1 Ω	1 A	$2,4 \cdot 10^{-3} \text{ m}\Omega$	
	10 Ω	100 mA	$2,4 \cdot 10^{-3} \Omega$	

Annex of the certificate (Page 15/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
DC Resistance	$0 \Omega \leq R < 2 \Omega$		$2,1 \cdot 10^{-6} \cdot R + 4,1 \mu\Omega$	R: Resistance Direct measurement
DC Resistance	$2 \Omega \leq R < 20 \Omega$		$1,1 \cdot 10^{-6} \cdot R + 14 \mu\Omega$	
Standard,	$20 \Omega \leq R < 200 \Omega$		$9,0 \cdot 10^{-6} \cdot R + 48 \mu\Omega$	
Calibrator:	$0,2 \text{ k}\Omega \leq R < 2 \text{ k}\Omega$		$8,9 \cdot 10^{-6} \cdot R + 0,5 \text{ m}\Omega$	
Resistance,	$2 \text{ k}\Omega \leq R < 20 \text{ k}\Omega$		$1,5 \cdot 10^{-6} \cdot R + 2,7 \text{ m}\Omega$	
DC Current	$20 \text{ k}\Omega \leq R < 200 \text{ k}\Omega$		$1,1 \cdot 10^{-5} \cdot R + 34 \text{ m}\Omega$	
Shunt,	$0,2 \text{ M}\Omega \leq R < 2 \text{ M}\Omega$		$1,3 \cdot 10^{-5} \cdot R + 0,6 \Omega$	
Resistance Box	$2 \text{ M}\Omega \leq R < 20 \text{ M}\Omega$		$2,5 \cdot 10^{-5} \cdot R + 95 \Omega$	
	$20 \text{ M}\Omega \leq R < 200 \text{ M}\Omega$		$1,3 \cdot 10^{-4} \cdot R + 10,2 \text{ k}\Omega$	
	$0,2 \text{ G}\Omega \leq R < 2 \text{ G}\Omega$		$1,6 \cdot 10^{-3} \cdot R + 1,1 \text{ M}\Omega$	
	$2 \text{ G}\Omega \leq R < 20 \text{ G}\Omega$		$1,6 \cdot 10^{-3} \cdot R + 10,2 \text{ M}\Omega$	
Capacitance	$220,0 \text{ pF} \leq C < 399,9 \text{ pF}$	1 kHz	$4,1 \cdot 10^{-3} \cdot C + 7,7 \text{ pF}$	C: Capacitance
Capacitance	$0,4 \text{ nF} \leq C < 1,0999 \text{ nF}$	1 kHz	$4,0 \cdot 10^{-3} \cdot C + 8 \text{ pF}$	
Measuring	$1,1 \text{ nF} \leq C < 3,2999 \text{ nF}$	1 kHz	$4,0 \cdot 10^{-3} \cdot C + 8 \text{ pF}$	
Instruments,	$3,3 \text{ nF} \leq C < 10,999 \text{ nF}$	1 kHz	$2,1 \cdot 10^{-3} \cdot C + 8 \text{ pF}$	
LCR Meter	$11 \text{ nF} \leq C < 32,999 \text{ nF}$	1 kHz	$2,0 \cdot 10^{-3} \cdot C + 0,08 \text{ nF}$	
	$33 \text{ nF} \leq C < 109,99 \text{ nF}$	1 kHz	$2,1 \cdot 10^{-3} \cdot C + 0,07 \text{ nF}$	
	$110 \text{ nF} \leq C < 329,99 \text{ nF}$	1 kHz	$2,1 \cdot 10^{-3} \cdot C + 0,23 \text{ nF}$	
	$0,33 \mu\text{F} \leq C < 1,0999 \mu\text{F}$	1 kHz	$2,3 \cdot 10^{-3} \cdot C + 0,75 \text{ nF}$	
	$1,1 \mu\text{F} \leq C < 3,2999 \mu\text{F}$	1 kHz	$2,2 \cdot 10^{-3} \cdot C + 2,3 \text{ nF}$	
	$3,3 \mu\text{F} \leq C < 10,999 \mu\text{F}$	1 kHz	$2,5 \cdot 10^{-3} \cdot C + 7 \text{ nF}$	

Annex of the certificate (Page 16/38)

Accreditation Scope

 Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ.</p> <p>Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K</p> <p>Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Capacitance Capacitance Measuring Instruments, LCR Meter	$11 \mu\text{F} \leq C < 32,999 \mu\text{F}$	100 Hz	$3,5 \cdot 10^{-3} \cdot C + 0,023 \mu\text{F}$	C: Capacitance
	$33 \mu\text{F} \leq C < 109,99 \mu\text{F}$	100 Hz	$3,9 \cdot 10^{-3} \cdot C + 0,07 \mu\text{F}$	
	$110 \mu\text{F} \leq C < 329,99 \mu\text{F}$	100 Hz	$4,1 \cdot 10^{-3} \cdot C + 0,20 \mu\text{F}$	
TIME and FREQUENCY Frequency Sources, Frequency Generator	$1 \text{ Hz} \leq f \leq 3\text{GHz}$		$7,1 \cdot 10^{-6} \cdot f + 11 \text{ mHz}$	f: Frequency Direct Measurement
Frequency Frequency Measuring Instruments, Frequency Counter	$1 \text{ Hz} \leq f < 100 \text{ kHz}$		$2,4 \cdot 10^{-5} \cdot f + 4 \text{ mHz}$	f: Frequency Direct Measurement
	$100 \text{ kHz} \leq f \leq 6 \text{ GHz}$		$1,3 \cdot 10^{-7} \cdot f + 0,2 \text{ mHz}$	
Frequency Frequency Measuring Instruments, Frequency Counter	$10 \text{ Hz} \leq f < 120 \text{ Hz}$		$4,0 \cdot 10^{-6} \cdot f + 9 \text{ mHz}$	f: Frequency Direct measurement
	$120 \text{ Hz} \leq f < 1200 \text{ Hz}$		$2,8 \cdot 10^{-6} \cdot f + 85 \text{ mHz}$	with FLUKE 5500A
	$1,2 \text{ kHz} \leq f < 12 \text{ kHz}$		$1,5 \cdot 10^{-5} \cdot f + 50 \text{ mHz}$	
	$12 \text{ kHz} \leq f < 120 \text{ kHz}$		$1,9 \cdot 10^{-5} \cdot f + 10 \text{ mHz}$	
Time Interval Time Interval Measuring Instruments, Time Difference Measuring Instruments, Frequency Counter	4ns - 0,1 s		$1,0 \cdot 10^{-9} \text{ s}$	t: Time Comparison method
	$10 \text{ s} \leq t \leq 3600 \text{ s}$		$9,0 \cdot 10^{-5} \text{ s}$	

Annex of the certificate (Page 17/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Time Interval Time Interval Measuring Instruments, Time Difference Measuring Instruments, Stopwatch (chronometer), Timer	$10 \mu s \leq t \leq 36000 s$		0,03 s	t: Time
Radio Frequency (RF) Power Absolute RF Power RF Power Source RF Power Source, Reference Power Output of RF Power Measuring Instruments, Signal Generator etc.	100 pW - 10 μ W (between -70 dBm and -30 dBm) 3 μ W - 100 mW (between -25 dBm and +20 dBm)	0,05 - 18 GHz 0,01 - 18 GHz	% 9 · P % 6 · P	P: Power

Annex of the certificate (Page 18/38)


Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Radio Frequency (RF) Power Absolute RF Power RF Power Meter Power measurement functions of Spectrum Analyzer, Test Set, Site Analyzer etc.	100 pW - 10 μW (between -70 dBm and -30 dBm) 3 μW - 100 mW (between -25 dBm and +20 dBm)	0,05 - 6 GHz 0,01 - 6 GHz	% 11 · P % 11 · P	P: Power
Signal and Pulse Characteristics Horizontal Deflection (Time) Oscilloscope	$2 \text{ ns} \leq t < 5 \text{ s}$	$> 1 \text{ V pk}$	% 1,0	t: Time Sawtooth Pulse
Signal and Pulse Characteristics Vertical Deflection (Gain) Oscilloscope	$5 \text{ mV} \leq U \leq 5 \text{ V}$ $5 \text{ mV} \leq U \leq 33 \text{ V}$ $5 \text{ mV} \leq U \leq 2,0 \text{ V}$	10 Hz - 10 kHz DC 1 MΩ DC 50 Ω	% 1,2 % 1,0 % 1,2	U: Voltage Square wave Direct measurement
Signal and Pulse Characteristics Rise Time Oscilloscope	$1,4 \text{ ns} - 7 \text{ s}$		% 3,0	
Signal and Pulse Characteristics Band Width Oscilloscope	$BW \leq 250 \text{ MHz}$ $250 \text{ MHz} < BW \leq 1\text{GHz}$		% 3,0 % 2,8	BW: Band width

Annex of the certificate (Page 19/38)


Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Time and Frequency Frequency Revolution source, Centrifuge, Stroboscope	15 rpm < w ≤ 99,999 rpm 100 rpm < w ≤ 999,99 rpm 1000 rpm < w ≤ 9999,9 rpm 10000 rpm < w ≤ 99999 rpm		$1,6 \cdot 10^{-4} \cdot w + 0,012$ rpm $2,6 \cdot 10^{-4} \cdot w + 0,002$ rpm $2,6 \cdot 10^{-4} \cdot w + 0,008$ rpm $4,9 \cdot 10^{-5} \cdot w + 2,2$ rpm	w: Revolution
Time and Frequency Frequency Optical Tachometer	15 rpm < w ≤ 100000 rpm		$5,6 \cdot 10^{-6} \cdot w + 0,06$ rpm	w: Revolution
TEMPERATURE Direct Reading Thermometer Resistance Thermometer Thermocouple Thermometer	-30 °C ≤ T ≤ +200 °C		0,06 °C	T: Temperature Calibration in liquid baths including ice point uncertainty Comparison method with standard platinum resistance thermometer
Direct Reading Thermometer Resistance Thermometer Thermocouple Thermometer	-90 °C ≤ T ≤ -25 °C -25 °C < T ≤ +50 °C +50 °C < T ≤ +650 °C		0,11 °C 0,08 °C 0,11 °C	T: Temperature Calibration on-site(field) by dry block calibrators Comparison method with standard platinum resistance thermometer

Annex of the certificate (Page 20/38)


Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Direct Reading Thermometer Resistance Thermometer Thermocouple Thermometer	+650 °C < T ≤ +1150 °C		2,1 °C	T: Temperature Calibration on-site(field) by dry block calibrators Comparison method with standard type-S thermocouple
Liquid-In-Glass Thermometer	-30 °C ≤ T ≤ +200 °C		0,06 °C	T: Temperature Calibration in liquid baths including ice point uncertainty Comparison method
Controlled Enclosures (Temperature Distribution) Test Chamber, Incubator, Oven, Steriliser, Deep Freeze, Fridge, Cold Chamber Room, etc.	-70 °C ≤ T ≤ -30 °C -30 °C < T ≤ +100 °C +100 °C < T ≤ +200 °C +200 °C < T ≤ +350 °C		0,7 °C 0,45 °C 0,7 °C 1,2 °C	T: Temperature On-site(field) calibration, temperature distribution of test chamber according to EN 60068-3-5, EN 60068-3-6, EN 60068-3-7, EN 60068-3-11, EURAMET cg-20, DKD R5-7
Controlled Enclosures (Temperature and/or Humidity Distribution) Climatic Chamber, Stability Chamber, Humidity Chamber, Stability Room etc.	70 °C ≤ T ≤ -30 °C -30 °C < T ≤ +100 °C +100 °C < T ≤ +200 °C +200 °C < T ≤ +350 °C 10 %rh ≤ RH ≤ 50 %rh 50 %rh < RH ≤ 80 %rh 80 %rh < RH ≤ 95 %rh	10 °C.. 50°C 10 °C.. 50°C 10 °C.. 50°C	0,7 °C 0,45 °C 0,7 °C 1,2 °C 0,60 %rh 0,85 %rh 0,95 %rh	T: Temperature , RH: Humidity On-site(field) calibration, temperature and/or humidity distribution of test chamber according to EN 60068-3-5, EN 60068-3-6 EN 60068-3-7, EN 60068-3-11 EURAMET cg-20 DKD R5-7

Annex of the certificate (Page 21/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Controlled Enclosures (Temperature Distribution) Furnace (Ashing Furnace)	+200 °C ≤ T ≤ +1150 °C		1,1 °C	T: Temperature On-site(Field) calibration
Controlled Enclosures (Temperature Distribution) Liquid Baths (Calibration Baths)	-90 °C ≤ T ≤ -30 °C -30 °C < T ≤ +200 °C		0,13 °C 0,08 °C	T: Temperature Determination of temperature distribution of bath On-site(field) calibration
Dry Block Calibrators	-90 °C ≤ T ≤ +140 °C +140 °C < T ≤ +650 °C +650 °C < T ≤ +1150 °C		0,09 °C 0,1 °C 1,7 °C	T: Temperature In accordance with Euramet cg-13 / v:02 By standard platinum resistance thermometer for single or poly holed types By standard S-type thermocouple for single or poly holed types

Annex of the certificate (Page 22/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Temperature Indicator		Cold junction "Off"		7: Temperature
PT100	-200 °C ≤ T ≤ +650 °C	At permanent laboratory facility or on-site(field)	0,4 °C	By electrical simulation method using electrical references in accordance with IEC 584-1, DIN 43710, EURAMET cg-8 v2.1, EURAMET cg-11
Type B Thermocouple	+600 °C ≤ T ≤ +1800 °C		1,6 °C	
Type R Thermocouple	0 °C ≤ T ≤ +1700 °C		1,6 °C	
Type S Thermocouple	0 °C ≤ T ≤ +1700 °C		1,5 °C	
Type J Thermocouple	-200 °C ≤ T ≤ +1200 °C		0,5 °C	
Type K Thermocouple	-200 °C ≤ T ≤ +1300 °C		0,7 °C	
Type T Thermocouple	-200 °C ≤ T ≤ +400 °C		0,9 °C	

Annex of the certificate (Page 23/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Temperature Calibrator PT100 Type B Thermocouple Type R Thermocouple Type S Thermocouple Type J Thermocouple Type K Thermocouple Type T Thermocouple	-200 °C ≤ T ≤ +650 °C +600 °C ≤ T ≤ +1800 °C 0 °C ≤ T ≤ +1700 °C 0 °C ≤ T ≤ +1700 °C -200 °C ≤ T ≤ +1200 °C -200 °C ≤ T ≤ +1300 °C -200 °C ≤ T ≤ +400 °C	Cold junction "Off" At permanent laboratory facility	0,2 °C 0,9 °C 0,9 °C 0,8 °C 0,4 °C 0,6 °C 0,8 °C	T: Temperature By electrical simulation method using electrical references in accordance with IEC 584-1, DIN 43710, EURAMET cg-8 v2.1
RADIATION THERMOMETRY Pyrometer, Thermal Camera, IR Thermometer, IR Ear Thermometer	-30 °C ≤ T ≤ +0 °C +0 °C < T ≤ +100 °C +100 °C < T ≤ +300 °C +300 °C < T ≤ +500 °C	For Emissivity = %95	2,25 °C 2,20 °C 3,60 °C 5,45 °C	T: Temperature Comparison method Calibration with blackbody

Annex of the certificate (Page 24/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
HUMIDITY Direct Reading Thermometer, Direct Reading Thermo Hygrometer, Direct Reading Hygrometer, Hygrometer, Relative Humidity Meter, Dew Point Temperature Measuring Instrument	-70 °C ≤ T ≤ -30 °C -30 °C < T ≤ +10 °C +10 °C < T ≤ +60 °C +60 °C < T ≤ +180 °C 10 %rh ≤ RH ≤ 50 %rh 50 %rh < RH ≤ 80 %rh 80 %rh < RH ≤ 95 %rh	Inside Temperature and/or Humidity Chamber 10 °C.. 50°C 10 °C.. 50°C 10 °C.. 50°C	0,8 °C 0,27 °C 0,17 °C 0,55 °C 0,65 %rh 0,85 %rh 1,05 %rh	T: Temperature RH: Humidity Comparison method
Dew Point Temperature Dew Point Temperature Measuring Instrument	+10°C ≤ DP ≤ +60°C	+10°C..+60°C	0,60 °C	DP : Dew Point Comparison
PRESSURE Relative Pressure Analogue Pressure Gauge, Digital Pressure Gauge, Differential Pressure Gauge, Presssure Transmitter, Pressure Switch	-99 kPa ≤ p ≤ - 10 kPa 5 Pa ≤ p ≤ 100 Pa 100 Pa < p ≤ 3750 Pa 3,75 kPa < p ≤ 10 kPa 10 kPa < p ≤ 34,5 kPa 34,5 kPa < p ≤ 3,5 MPa 200 kPa ≤ p ≤ 70 MPa	Pneumatic Pneumatic Pneumatic Pneumatic Pneumatic Pneumatic Hydraulic	300 Pa 0,8 Pa $6,0 \cdot 10^{-3} \cdot p + 1,8 \text{ Pa}$ $6,0 \cdot 10^{-3} \cdot p + 5 \text{ Pa}$ $7,0 \cdot 10^{-3} \cdot p + 20 \text{ Pa}$ $6,0 \cdot 10^{-4} \cdot p + 200 \text{ Pa}$ $6,0 \cdot 10^{-4} \cdot p + 400 \text{ Pa}$	Calibrations are performed according to basic and standard calibration procedure defined in Euramet cg-17/v4.0 "Guidelines on the Calibration of Electromechanical and Mechanical Manometers" p: Pressure (Pa)

Annex of the certificate (Page 25/38)

Accreditation Scope

 Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ.</p> <p>Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K</p> <p>Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Absolute Pressure Analogue Pressure Gauge, Digital Pressure Gauge, Pressure Transmitter	$1,5 \text{ kPa} \leq p \leq 1,6 \text{ MPa}$ $1,6 \text{ MPa} < p \leq 3,5 \text{ MPa}$	Pneumatic	$5,0 \cdot 10^{-4} \cdot p + 300 \text{ Pa}$ $6,0 \cdot 10^{-4} \cdot p + 300 \text{ Pa}$	Calibrations are performed according to basic and standard calibration procedure defined in Euramet cg-17/v4.0 "Guidelines on the Calibration of Electromechanical and Mechanical Manometers" <i>p</i> : Pressure (Pa)
TORQUE Hand Torque Tools, Reference Torque Wrenches	$0,2 \text{ N}\cdot\text{m} \leq M \leq 1 \text{ N}\cdot\text{m}$ $1 \text{ N}\cdot\text{m} < M \leq 1000 \text{ N}\cdot\text{m}$	Clockwise and Counter-Clockwise direction	% 1,5 % 1,0	<i>M</i> : Torque Calibration procedure in accordance with TS EN ISO 6789-1 and TS EN ISO 6789-2
HARDNESS Shore Hardness Tester, Durometer	$0 \text{ Shore A} \leq \text{Shore} \leq 100 \text{ Shore A}$ $10 \text{ Shore D} \leq \text{Shore} \leq 100 \text{ Shore D}$	Depth of Indentation Spring Force Shore A Spring Force Shore D Diameter Angle	$2 \mu\text{m}$ $7,5 \text{ mN}$ $44,5 \text{ mN}$ $(2 + 17 \cdot D) \mu\text{m}$ $3'$ 1 Shore A 1 Shore D	<i>Shore</i> : Hardness Value <i>D</i> : Diameter, m Calibration procedure in accordance with ISO 48-9
DIMENSION Caliper (Outside, inside, depth, step measurements)	$L \leq 1500 \text{ mm}$	Inside and Outside Measurements Depth Measurements Step	$(10 + 9 \cdot L) \mu\text{m}$	<i>L</i> : Length, m Calibration procedure in accordance with DAkkS-DKD-R 4-3 Blatt 09.1

Annex of the certificate (Page 26/38)

Accreditation Scope

 <p>TÜRKAK Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Depth Caliper	$L \leq 1000$ mm		$(11 + 8 \cdot L) \mu\text{m}$	L: Length, m Calibration procedure in accordance with DAKKS-DKD-R 4-3 Blatt 09.2
Outside Micrometer	$L \leq 1000$ mm		$(1,5 + 13 \cdot L) \mu\text{m}$	L: Length, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.10.1
Depth Micrometer	$L \leq 300$ mm		$(1,5 + 12 \cdot L) \mu\text{m}$	L: Length, m Calibration procedure in accordance with VDI/VDE/DGQ 2618 Bl.10.5
Height Measuring Instrument (Height Gauge)	$L \leq 1000$ mm		$(4 + 11 \cdot L) \mu\text{m}$	L: Length, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.9.3
Dial Gauge, Digital Dial Gauge (Comparator)	$L \leq 100$ mm		$(0,7 + 40 \cdot L) \mu\text{m}$	L: Length, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.11.1
Lever Gauge	$L \leq 2$ mm		0,7 μm	L: Length Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.11.3

Annex of the certificate (Page 27/38)


Accreditation Scope

 <p>TÜRKAK Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Dial Indicator, Digital Indicator	$L \leq 2 \text{ mm}$		0,7 μm	L: Length Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.11.2
Pin Gauge (Thread measuring wires)	$0,1 \text{ mm} \leq D \leq 20 \text{ mm}$		0,5 μm	D: Diameter Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.04.02
Micrometer Setting Gauge	$25 \text{ mm} \leq L \leq 675 \text{ mm}$		$(0,5 + 9 \cdot L) \mu\text{m}$	L: Length, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.04.04
Thickness Gauge, Feeler Gauge	$0,03 \text{ mm} \leq L \leq 2 \text{ mm}$		0,5 μm	L: Gauge dimension Calibration procedure in accordance with DIN 2275
Radius Gauge, Fillet Gauge, Radius Template	$0,03 \text{ mm} \leq r \leq 100 \text{ mm}$		9 μm	r: Radius Optical Measurement Method
Test Sieve	$0,02 \text{ mm} \leq L \leq 125 \text{ mm}$		$(5 + 32 \cdot L) \mu\text{m}$	L: Measured Opening, m Calibration procedure in accordance with TS ISO 3310-1
Bevel Protractor, Protractor	$\alpha \leq 360^\circ$	Measuring arm up to 400mm	0,6' (Angle) 3,5 μm (Flatness) 3,8 μm (Parallelism)	α : Angle Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.07.2

Annex of the certificate (Page 28/38)

Accreditation Scope

 <p>TÜRKAK Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Square 90°	L ≤ 400 mm (Squareness) L ≤ 500 mm (Straightness/Flatness)		3 µm (0,6 + 1,2 · L) µm	L: Square Length, m Calibration procedure in accordance with VDI/VDE/DGQ 2618 Bl.7.1
Knife Edge	L ≤ 500 mm (Straightness/Flatness)		(0,6 + 1,2 · L) µm	L: Knife Edge Length, m Calibration procedure in accordance with VDI/VDE/DGQ 2618 Bl.5.2
Gauge Block	0,5 mm ≤ L ≤ 100 mm	Central Length Deviation	(0,1 + 0,6 · L) µm (steel) (0,1 + 0,6 · L) µm (ceramic) (0,1 + 2,2 · L) µm (tungsten carbide)	L:Gauge Block Length, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.03.01
		Variation	0,06 µm	
Gauge Block	0,1 mm ≤ L ≤ 500 mm	Steel, Ceramic and Tungsten Carbide	(1 + 9 · L) µm	L:Gauge Block Length, m Measuring of central length deviation by comparison method. Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.03.01.

Annex of the certificate (Page 29/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Steel Rule	$L \leq 150 \text{ mm}$ $150 \text{ mm} \leq L \leq 2000 \text{ mm}$ $L \leq 3000 \text{ mm}$	<p>Calibration by measuring microscope</p> <p>Calibration by measuring microscope</p> <p>Comparison with reference steel rule</p>	$25 \mu\text{m}$ $(25 + 125 \cdot L) \mu\text{m}$ $(250 + 100 \cdot L) \mu\text{m}$	<p>L: Length, m Calibration procedure in accordance with DIN 866</p>
Tape measure (surveying, workshop), Circometer (Pi tape), Grade rod, Telescopic measure, Flexible tape measure, Wooden measure, Plastic measure, Folding measure	$L \leq 100 \text{ m}$		$(770 + 40 \cdot L) \mu\text{m}$	L: Length, m Calibration procedure in accordance with TS 9505 ve OIML R35-1
Profile Projector, Measuring Microscope	$L \leq 300 \text{ mm}$ (Axial measurements) $\alpha \leq 180^\circ$ (Angle measurements) $r \leq 10 \text{ mm}$ (Radius measurements)		$(1,7 + 9 \cdot L) \mu\text{m}$ $1,5'$ $10 \mu\text{m}$	<p>L: Length, m α: Angle r: Radius Comparison method with glass scale</p>
Inside Cylindrical Ring Gauge (Reference, go & nogo, etc.)	$1 \text{ mm} \leq D \leq 150 \text{ mm}$ $150 \text{ mm} \leq D \leq 250 \text{ mm}$		$(0,6 + 6 \cdot D) \mu\text{m}$ $(1,6 + 4 \cdot D) \mu\text{m}$	<p>D: Diameter, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.4.01</p>

Annex of the certificate (Page 30/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Outside Cylindrical Plug Gauge (Reference, go & nogo, etc.)	$1 \text{ mm} \leq D \leq 600 \text{ mm}$		$(0,5 + 12 \cdot D) \mu\text{m}$	D: Diameter, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.4.01
Gap Gauge	$1 \text{ mm} \leq L \leq 150 \text{ mm}$ $150 \text{ mm} \leq L \leq 250 \text{ mm}$		$(0,6 + 6 \cdot L) \mu\text{m}$ $(1,6 + 4 \cdot L) \mu\text{m}$	L: Length, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.04.07
Cylindrical Thread Ring Gauge	$3 \text{ mm} \leq D \leq 100 \text{ mm}$	Pitch: 0,5-4,5 mm	$(2 + 40 \cdot D) \mu\text{m}$	D: Pitch Diameter, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.04.09
Cylindrical Thread Plug Gauge	$1 \text{ mm} \leq D \leq 200 \text{ mm}$	Pitch: 0,3-5,5 mm	$(3 + 45 \cdot D) \mu\text{m}$	D: Pitch Diameter, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.04.08
Tapered Thread Plug Gauge	$1 \text{ mm} \leq D \leq 200 \text{ mm}$	Pitch: 0,3-5,5 mm	6 μm	D: Pitch Diameter Calibration procedure in accordance with ASME B1.20.5
Tapered Thread Ring Gauge	$3 \text{ mm} \leq D \leq 100 \text{ mm}$	Pitch: 0,5-4,5 mm	4 μm	D: Pitch Diameter Calibration procedure in accordance with ASME B1.20.5
Thread Pitch Gauge	$L \leq 7 \text{ mm}$		9 μm	L: Measured Pitch Optical Measurement Method

Annex of the certificate (Page 31/38)


Accreditation Scope

 <p>TÜRKAK</p> <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ.</p> <p>Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K</p> <p>Revision Nr: 015 Date: 19.11.2019</p>
---	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Coating Thickness Gauge	$L \leq 6000 \mu\text{m}$		$(1,1 + 50 \cdot L) \mu\text{m}$	L: Measured Thickness, m Comparison Method
Coating Thickness Standard (Shim, Foil)	$7 \mu\text{m} \leq L \leq 8000 \mu\text{m}$		$1,0 \mu\text{m}$	L: Measured Thickness Comparison Method
Thickness Gauge (Dial, digital)	$L \leq 50 \text{ mm}$		$1,5 \mu\text{m}$	L: Measured Thickness Comparison with block gauges
Caliper Gauge (External)	$L \leq 200 \text{ mm}$		$(1,5 + 10 \cdot L) \mu\text{m}$	L: Length, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.12.1
Caliper Gauge (Internal)	$3 \text{ mm} \leq L \leq 200 \text{ mm}$		$(1,5 + 10 \cdot L) \mu\text{m}$	L: Length, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.13.1
Glass Linear Scale, Stage Micrometer	$L \leq 150 \text{ mm}$		$(4 + 35 \cdot L) \mu\text{m}$	L: Length Optical Measurement Method
Internal Micrometer (2-Point Contact)	$10 \text{ mm} \leq L \leq 600 \text{ mm}$		$(1,2 + 11 \cdot L) \mu\text{m}$	L: Length, m Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.10.7
Internal Micrometer (2-Line Contact)	$5 \text{ mm} \leq L \leq 200 \text{ mm}$		$(1,5 + 10 \cdot L) \mu\text{m}$	L: Length Comparison Method

Annex of the certificate (Page 32/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Internal Micrometer (3-Line Contact)	$3 \text{ mm} \leq D \leq 100 \text{ mm}$		2,5 μm	D: Diameter Calibration procedure in accordance with Dakks-DKD-R 4.3 Bl.10.8
Level Inclinometer	$L \leq 200 \text{ mm}$ $L \leq 800 \text{ mm}$ $L \leq 200 \text{ mm}$		0,04 mm/m 0,3 mm/m 0,4'	L: Base Length Calibration procedure in accordance with DIN 877
Optical Flat, Optical Parallel	$D \leq 45 \text{ mm}$		0,2 μm (Flatness) 0,06 μm (Parallelism)	D: Diameter Calibration procedure in accordance with VDI/VDE/DGQ 2618 Bl.6.1
V-Block	$L \leq 300 \text{ mm}$	Squareness Parallelism Flatness	2,0 μm 0,9 μm 0,6 μm	L: Base Length Calibration procedure in accordance with DIN 2274
Ultrasonic Thickness Gauge	$L \leq 100 \text{ mm}$		0,174 mm	L: Length Measurement with steel gauge blocks
Wet Film Thickness Gauge	$d \leq 1800 \mu\text{m}$		4 μm	d: Depth Calibration procedure in accordance with ISO 2808
Grindometer (Fineness of Grind Gauge)	$d \leq 1800 \mu\text{m}$		4 μm	d: Depth Calibration procedure in accordance with EN ISO 1524

Annex of the certificate (Page 33/38)

Accreditation Scope

 <p>TÜRKAK Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Cross Cut Tester (Cross Cut Cutting Tool)	L: 5 - 10 mm	Distance Angle	9 µm 18'	L : Distance of cutting blades Calibration procedure in accordance with EN ISO 2409
Gap Wedge Gauge (Gap Feeler Gauge, Taper Gauge)	L ≤ 100 mm		14 µm	L: Length Optical Measurement Method
Laser Diameter Gauge	0,5 mm ≤ D ≤ 20 mm		2,5 µm	D: Diameter Comparison with pin gauges
Micrometer Head	L ≤ 100 mm		(0,7 + 40 · L) µm	L: Length Calibration procedure in accordance with DIN 863-2
Laser Distance Meter	L ≤ 10 m 10 m < L ≤ 50 m		5 mm 12 mm	L: Length Comparison method with reference laser distance meter
Optical Level	± 1°	Levelling Error	9,0" (2,8 mgon)	Measurement of angular errors with optical collimator system
Theodolite, Total Station	± 1° ± 1° 0,25°	Levelling Error Horizontal/Vertical Collimation and Horizontal Tilting Errors Plummet Error	5,1" (1,6 mgon) 5,6" (1,7 mgon) 17" (5,3 mgon)	Measurement of angular errors with optical collimator system. Measurement of plummet error with reference scale.

Annex of the certificate (Page 34/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Non-Automatic Weighing Instruments Weighing Scale (Balance)	$1 \text{ mg} \leq m \leq 2200 \text{ g}$	By class E2 weights	$1 \cdot 10^{-6}$	<i>m</i> : Weight Calibration procedure in accordance with EURAMET/cg-18/v.04
	$1 \text{ g} \leq m \leq 40 \text{ kg}$	By class F1 weights	$1 \cdot 10^{-5}$	
	$5 \text{ kg} \leq m \leq 3000 \text{ kg}$	By class M1 weights	$1 \cdot 10^{-4}$	
	$10 \text{ kg} \leq m \leq 15000 \text{ kg}$	By substitute weights	$1,5 \cdot 10^{-4}$	
WEIGHTS (STANDARD WEIGHTS) Weight Class E2 Weights Conventional Mass Value	$1 \text{ mg} \leq m \leq 20 \text{ mg}$ 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg	Air	0,003 mg 0,004 mg 0,005 mg 0,006 mg 0,008 mg 0,010 mg 0,012 mg 0,016 mg 0,020 mg 0,025 mg 0,03 mg 0,05 mg 0,10 mg 0,25 mg 0,5 mg	<i>m</i> : Nominal weight Calibration procedure in accordance with OIML R111-1

Annex of the certificate (Page 35/38)


Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Weight Class F1 Weights Conventional Mass Value	$1 \text{ mg} \leq m \leq 5 \text{ mg}$ 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg	Air	0,006 mg 0,008 mg 0,010 mg 0,012 mg 0,016 mg 0,020 mg 0,025 mg 0,03 mg 0,04 mg 0,05 mg 0,06 mg 0,08 mg 0,10 mg 0,16 mg 0,3 mg 0,8 mg 1,6 mg 3,0 mg 8,0 mg 16 mg	<i>m</i> : Nominal weight Calibration procedure in accordance with OIML R111-1
Weight Class F2 Weights Conventional Mass Value	$1 \text{ mg} \leq m \leq 5 \text{ mg}$ 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg	Air	0,020 mg 0,025 mg 0,03 mg 0,04 mg 0,05 mg 0,06 mg 0,08 mg 0,10 mg 0,12 mg 0,16 mg 0,20 mg 0,25 mg 0,3 mg 0,5 mg 1,0 mg 2,5 mg 5,0 mg 10 mg 25 mg 50 mg	<i>m</i> : Nominal weight Calibration procedure in accordance with OIML R111-1

Annex of the certificate (Page 36/38)

Accreditation Scope

 <p>Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K</p>	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	---

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Weight Class M1 Weights Conventional Mass Value	$1 \text{ mg} \leq m \leq 5 \text{ mg}$ 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg	Air	0,06 mg 0,08 mg 0,10 mg 0,12 mg 0,16 mg 0,20 mg 0,25 mg 0,3 mg 0,4 mg 0,5 mg 0,6 mg 0,8 mg 1,0 mg 1,6 mg 3,0 mg 8,0 mg 16 mg 30 mg 80 mg 160 mg 300 mg	<i>m</i> : Nominal weight Calibration procedure in accordance with OIML R111-1
Weight Class M2 Weights Conventional Mass Value	100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg	Air	0,5 mg 0,6 mg 0,8 mg 1,0 mg 1,2 mg 1,6 mg 2,0 mg 2,5 mg 3,0 mg 5,0 mg 10 mg 25 mg 50 mg 100 mg 250 mg 500 mg 1000 mg	Calibration procedure in accordance with OIML R111-1

Annex of the certificate (Page 37/38)

Accreditation Scope

 Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ.</p> <p>Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K</p> <p>Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Weight Class M3 Weights Conventional Mass Value	1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg	Air	3,0 mg 4,0 mg 5,0 mg 6,0 mg 8,0 mg 10 mg 16 mg 30 mg 80 mg 160 mg 300 mg 800 mg 1600 mg 3000 mg	Calibration procedure in accordance with OIML R111-1
Weight Non-standard weights Conventional Mass Value	1 g 1 g < m ≤ 2 g 2 g < m ≤ 5 g 5 g < m ≤ 10 g 10 g < m ≤ 20 g 20 g < m ≤ 50 g 50 g < m ≤ 100 g 100 g < m ≤ 200 g 200 g < m ≤ 500 g 500 g < m ≤ 1 kg 1 kg < m ≤ 2 kg 2 kg < m ≤ 5 kg 5 kg < m ≤ 10 kg 10 kg < m ≤ 20 kg 20 kg < m ≤ 30 kg	Air	0,3 mg 0,4 mg 0,5 mg 0,6 mg 0,8 mg 1,0 mg 1,6 mg 3,0 mg 8,0 mg 16 mg 30 mg 80 mg 160 mg 1000 mg 2500 mg	m: Nominal weight Calibration procedure in accordance with OIML R111-1
FLUID FLOW Volumetric Flow Rate of Gases Flow meter, Rotameter (Variable Area Flowmeter), Gas Meter	0,05 l/min ≤ Q ≤ 0,5 l/min 0,5 l/min < Q < 5 l/min 5 l/min ≤ Q < 10 l/min 10 l/min ≤ Q ≤ 100 l/min	(0 - 6) barg air	%5 %3,5 %1,8 %0,5	Q: Measured flow Comparison with reference flow meter

Annex of the certificate (Page 38/38)

Accreditation Scope

 Kalibrasyon TS EN ISO/IEC 17025 AB-0016-K	<p>METKAL ÖLÇÜ VE TEST SİSTEMLERİ SAN. VE TİC. LTD. ŞTİ. Kalibrasyon Laboratuvarı</p> <p>Accreditation Nr: AB-0016-K Revision Nr: 015 Date: 19.11.2019</p>
--	--

Measured Quantity/Calibrated Items	Range	Measurement Conditions	Calibration and Measurement Capability (k=2)	Remarks/ Calibration Method
Volumetric Flow Rate of Gases Balometer (Air Flow Hood)	$50 \text{ m}^3/\text{h} \leq Q < 360 \text{ m}^3/\text{h}$ $360 \text{ m}^3/\text{h} \leq Q < 3000 \text{ m}^3/\text{h}$	Air	9 m ³ /h %2,7	Q: Measured flow Comparison with reference anemometer inside the wind tunnel
Velocity of Air Anemometer (Hot wire, pitot, vane type)	$0,1 \text{ m/s} \leq V \leq 1,5 \text{ m/s}$ $1,5 \text{ m/s} < V \leq 3,0 \text{ m/s}$ $3,0 \text{ m/s} < V \leq 40 \text{ m/s}$	(-0,02 - 0) barg dry air	0,03 m/s 0,04 m/s %1,5	V: Velocity of air Comparison with reference anemometer inside the wind tunnel

End of Scope

G. Banu MÜDERRİSOĞLU
Secretary General